

# Declaration of Conformity



**Type of equipment:** Digital Breathalyzer Alcoscent  
**Type designation /model:** DA-7100 , FCA7000, Element, DA-8500, DA-8500E and Q7  
**Manufacturer:** DA Tech Co., Ltd.

**In accordance with the following Directives:**

2014/30/EU The Electromagnetic Compatibility Directive

**The following harmonized European standards and technical specifications have been applied:**

|  |  |
|--|--|
| EN 55011:2016 +A1:2017<br>+A2:2021     | Industrial, scientific and medical equipment. Radio-frequency disturbance characteristics. Limits and methods of measurement   |
| EN IEC 61326-1:2021                    | Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements   |
| EN IEC 61326-2-2:2021                  | Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems |
| EN IEC 61326-2-3:2021                  | Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 2-3: Particular requirements - Test configuration, operational conditions and performance criteria for transducers with integrated or remote signal conditioning                      |
| EN 61000-4-3:2006 +A1:2008<br>+A2:2010 | Radiated, radio-frequency, electromagnetic field immunity test   |

**Test report issued by:**  
EMC: CTK Co., Ltd.

**The CE Marking on the products and/or their packaging signifies that DA Tech Co., Ltd. holds the reference technical file available to the European Union authorities.**

**Place and date of issue:** 13, Seounsandan-ro 4-gil, Gyeyang-gu, Incheon, Korea  
/ 2022-11-21

**Authorized Signatory:** Name: Kwang-hee, Park

Title: CEO

Signature: