



Product Service

Attestation of Conformity

No. E8A 17 06 55975 067

Holder of Certificate: **SENTECH KOREA CORP.**

21-6, Jimokro 75-gil (Sinchon-dong)
Paju-si, Gyeonggi-do 10880
REPUBLIC OF KOREA

Name of Object: **Electrical equ. for measurement, control and laboratory use (Breathalyzer)**

Model(s): **AL7000**

Description of Object: Rated Input Voltage: 3 V d.c.
EMC Classification: Class B equipment according to EN 61326-1

Tested according to: EN 61326-1:2013
EN 61326-2-3:2013
EN 55011:2009/A1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: TR-W1706-008



Date, 2017-06-19 (Byung-Soo Kang)

CE After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. That declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.