

**CE DECLARATION OF CONFORMITY****Plum QuickFix Plaster****Plaster for first aid**

We, the manufacturer, of the products covered by this declaration, hereby declare, that the plasters listed below, meet the requirements in the Council Directive 93/42/EEC of 14 June 1993, the Directive 2007/47/EC and its relevant transposition into national laws of the member states into which we place the device. Plum A/S is responsible for the Declaration of Conformity concerning the medical devices listed below:

<b>Art No</b>	<b>Art Name</b>
5511	QuickFix water resistant
5512	QuickFix elastic
5508	QuickFix elastic long
5504	QuickFix elastic Mini
5518	QuickFix elastic Micro
5513	QuickFix detectable
5509	QuickFix detectable long
5515	QuickFix alu

With the intended use:

Plum QuickFix plasters are be used as a first aid product as mechanical barrier and for absorption of exudates.

We guarantee and declare that:

1. The device is a Class I device according to Annex IX, rule 4.
2. The planning and manufacture procedures are conducted in conformity with the Company Quality System, in compliance with the provisions of Annex VII combined with Annex V to the Directive.

Assens, 2019-10-30  
Plum A/S



Lise S Bollen  
Quality Manager