

**FLUKE®****EU DECLARATION OF CONFORMITY**Product Identification:**Fluke 302+, 303, 305, 323, 324, 324+, 325 - Clamp Meter**

First declared: 2012

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU, RoHS**Directive 2014/30/EU, Electromagnetic Compatibility (EMC)****Directive 2014/35/EU, Low Voltage (LVD)**Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances
EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
EN 61010-2-032: 2002 Particular requirements for hand-held and hand manipulated current sensors
EN 61010-2-033: 2012 Particular requirements for Hand-Held Multimeters for measuring Mains voltage

Special Conditions:

None

Manufacturer:

Fluke
6920 Seaway Boulevard M/S 143F
Everett, WA 98203, USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20160526093303 - This declaration supersedes all previous declarations for this product.

26 May 2016