

Declaration of Conformity

MANUFACTURER: i-SENS, Inc.
43, Banpo-daero 28-gil, Seocho-gu,
Seoul 06646, Korea

MANUFACTURING FACILITY: i-SENS Wonju Factory
94-1, Donghwagongdan-ro, Munmak-eup,
Wonju-si, Gangwon-do 26365, Korea

EUROPEAN REPRESENTATIVE: Medical Technology Promedt Consulting
GmbH
Altenhofstrasse 80,
66386 St. Ingbert, Germany

PRODUCT: Blood Glucose Monitoring System

Model: See List of Products

CLASSIFICATION: List B according to Annex II of IVDD

CONFORMITY ASSESSMENT ROUTE: IVDD ANNEX IV without section 4 and 6
Applied

We herewith declare under our sole responsibility of the manufacturer that the above-mentioned products meet the provision of the Council Directive 98/79/EC for in vitro diagnostic medical devices. All supporting documentation is retained at the premises of the manufacturer.

STANDARD APPLIED: See List of Applied Standards

NOTIFIED BODY: TÜV SÜD PRODUCT SERVICE GmbH
Ridlerstraße 65, 80339 Munich, Germany
(Notified Body Number 0123)

CERTIFICATE: V1 090700 0028 Rev. 00

START OF CE-MARKING: See List of Products

PLACE, DATE OF ISSUE: Seoul, 2019-09-16

SIGNATURE:



CEO
Geun Sig Cha

List of Products

Brand Name/Model

CareSens N Blood Glucose Monitoring System, Model GM505PAD/GM505PBD/GM505PCD

- CareSens N Blood Glucose Meter, EDMA: 21 06 01

- CareSens N Blood Glucose Test Strips, EDMA: 11 70 01 01 00

- CareSens Glucose Control Solutions, EDMA: 11 50 90 90 00

* **Start of CE Marking: 2014-10-21**

List of Applied Standards

| Document Number | Title of Document |
|------------------------------------|--|
| EN ISO 13485: 2016 | Medical devices - Quality management systems -Requirements for regulatory purposes |
| EN ISO 18113-1: 2011 | In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 1: Terms, definitions and general requirements |
| EN ISO 18113-4: 2011 | In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 4: In vitro diagnostic reagents for self-testing |
| EN ISO 18113-5: 2011 | In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 5: In vitro diagnostic instruments for self-testing |
| EN ISO 15223-1: 2016 | Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements |
| EN 13532: 2002 | General requirements for in vitro diagnostic medical devices for self-testing |
| EN 13612: 2002 | Performance evaluation of in vitro diagnostic medical devices |
| EN ISO 23640:2015 | In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents |
| EN ISO 14971: 2012 | Medical devices - Application of risk management to medical devices |
| EN ISO 15197: 2015 | In vitro diagnostic test systems - Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus |
| EN ISO 17511: 2003 | In vitro diagnostic medical devices - Measurements of quantities in samples of biological origin - Metrological traceability of values assigned to calibrators and control materials |
| EN 61010-1: 2001 Second edition | Safety requirements for electrical equipment for measurement, control, and laboratory use. General requirements |
| EN 61010-2-101: 2002 | Safety requirements for electrical equipment for measurement, control and laboratory use. Particular requirements for in vitro diagnostic (IVD) medical equipment |
| EN 61326-2-6: 2006 | Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. In vitro diagnostic (IVD) medical equipment |
| EN 60068-2-64: 2008 | Environmental testing. Tests. Test Fh. Vibration, broadband random and guidance |
| EN 62304: 2006 | Medical device software - Software life cycle processes |
| EN 62366: 2008 | Medical devices - Application of usability engineering to medical devices |