

**FLUKE**

## EU DECLARATION OF CONFORMITY

Product Identification:

First declared: 2011

**Fluke T90, T110, T130, T150, T110VDE, T130VDE, T150VDE - Electrical Testers**

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

**Directive 2011/65/EU, RoHS**

**Directive 2014/30/EU, Electromagnetic Compatibility (EMC)**

**Directive 2014/35/EU, Low Voltage (LVD)**

Standards Used:

EN 60581: 2012 Restriction of the use of hazardous substances

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General

EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment

EN 61243-3: 2014 / AC2015 Live working -Voltage detectors; Two-pole low-voltage

.....

Special Conditions:

None

Manufacturer:

Fluke

6920 Seaway Boulevard M/S 143F

Everett, WA 98203; USA

DC2019011105209

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC2019011105209 - This declaration supersedes all previous declarations for this product.

11 January 2019