

POWDERFREE

Peha-taft® classic

POWDERFREE



Χειροπυκνά γάντια, χωρίς σκόνη, από λάτεξ φυσικού καουτσούκ
 Nepudrované chirurgické rukavice z prírodného kaučukového latexu
 Nepudrované chirurgické rukavice vyrobené z prírodného latexu
 Bezpuďrowe rękawiczki chirurgiczne wykonane z lateksu z naturalnego kauczuku
 Formentes sebeszeti kesztyű természetes latexgumiból
 Неопудренные хирургические перчатки, изготовленные из натурального латекса
 Хирургични рќавици без талк, изработени от естествен каучук
 Nenapudrane kirurške rukavice od prirodnog gumenog lateksa
 Mănuși chirurgicale nepudrate, fabricate din latex natural
 Nepudrane kirurške rokavice iz naravnega lateksa
 Looduslikust lateksist puudrivabad kirurgilised kindad
 Chirurgines piršines be pudros, pagamintos iš natūralios gumos latekso
 Kirurgiskie cimdi bez pulvera, kas izgatavoti no dabiskā gumijas lateksa
 Pudertia operationshandskar av naturgummilatex
 Pulverfri kirurgiske handsker lavet af naturlig gummilatex
 Puuterittomat luonnonlateksista valmistetut leikkauskäsineet
 Pudertrie operasjonshansker laget av naturgummilatex
 天然乳膠無粉手術手套
 قفازات جراحية خالية من المساحيق مصنوعة من لاتكس المطاط الطبي

CE 0 1 2 3 (Medical Device) 2 7 7 7 (Personal Protective Equipment)



Waterproof personal protective gloves against dangerous chemicals and microorganisms (cat.III). This product is classed as Personal Protective Equipment (PPE) and has been shown to comply with the Harmonised European Standards: EN 420, EN 421, EN ISO 374-1, EN 374-2, EN 16523-1, ISO 16604, ASTM F1671.



Tested acc. to EN ISO 374-1, EN 16523-1, EN 374-4
 Sodium hydroxide 40%, Level 6; Degradation* -11.0%
 Hydrogen peroxide 30%, Level 6; Degradation* 10.1%
 Formaldehyde 37%, Level 6; Degradation* 2.6%
 EN ISO 374-1: permeation levels are based on
 breakthrough times as follows:
 Level 1 < 10 min Level 4 < 120 min
 Level 2 < 30 min Level 5 < 240 min
 Level 3 < 480 min Level 6 < 720 min



Tested acc. to EN ISO 374-5, ISO 16604, ASTM F1671



Tested acc. to EN 421
 against mechanical risks
 Radioactive contamination protection
 only, excluding clause 4.3
 Note: This glove does not protect

Notified Body responsible for
 certification and for ongoing
 conformance is SATRA.
 SATRA Technology Europe Ltd.
 Bracetown Business Park, Clonee
 Dublin 15, Ireland (Notified Body 2777)