



**FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND**  
**9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864**

## ***Declaration of Conformity***

**Extech Model: IR200**

**Description: Non-Contact Forehead InfraRed Thermometer**

**Date of Issue: 19-Nov-20**

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

**EMC Directive: 2007/47/EC**  
**Report Number: 50055216 001**  
**Report Date of Issue: 10/12/2016**

**Standards:**

IEC 60601-1-2:2014

**RoHS Directive: EU Directive 2015/863/EU (RoHS 3)**

The test reports show that the product fulfills the requirement in the EC EMC Directive and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive and RoHS Directive.

A handwritten signature in black ink that reads "Mark Sultzbach".

---

Mark Sultzbach / QA Manager