

# *Declaration of Conformity*

**MANUFACTURER:** i-SENS, Inc.  
43, Banpo-daero 28-gil, Seocho-gu,  
Seoul 06646, Korea

**MANUFACTURING FACILITY:** i-SENS Wonju Factory  
94-1, Donghwagongdan-ro, Munmak-eup,  
Wonju-si, Gangwon-do 26365, Korea

**EUROPEAN REPRESENTATIVE:** Medical Technology Promedt Consulting  
GmbH  
Altenhofstrasse 80,  
66386 St. Ingbert, Germany

**PRODUCT:** Blood Glucose Monitoring System

**Model:** See List of Products

**CLASSIFICATION:** List B according to Annex II of IVDD

**CONFORMITY ASSESSMENT ROUTE:** IVDD ANNEX IV without section 4 and 6  
Applied

We herewith declare under our sole responsibility of the manufacturer that the above-mentioned products meet the provision of the Council Directive 98/79/EC for in vitro diagnostic medical devices. All supporting documentation is retained at the premises of the manufacturer.

**STANDARD APPLIED:** See List of Applied Standards

**NOTIFIED BODY:** TÜV SÜD PRODUCT SERVICE GmbH  
Ridlerstraße 65, 80339 Munich, Germany  
(Notified Body Number 0123)

**CERTIFICATE:** V1 090700 0028 Rev. 00

**START OF CE-MARKING:** See List of Products

**PLACE, DATE OF ISSUE:** Seoul, 2019-09-16

**SIGNATURE:**



---

CEO  
Geun Sig Cha

## **List of Products**

**Brand Name/Model**

CareSens N Premier Blood Glucose Monitoring System, Model GM01AAA/GM01AAB

- CareSens N Premier Blood Glucose Meter, EDMA: 21 06 01
- CareSens N Blood Glucose Test Strips, EDMA: 11 70 01 01 00
- CareSens Glucose Control Solutions, EDMA: 11 50 90 90 00

**\* Start of CE Marking: 2015-08-03**

## **List of Applied Standards**

<b>Document Number</b>	<b>Title of Document</b>
EN ISO 13485: 2016	Medical devices - Quality management systems -Requirements for regulatory purposes
EN ISO 18113-1: 2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 1: Terms, definitions and general requirements
EN ISO 18113-4: 2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 4: In vitro diagnostic reagents for self-testing
EN ISO 18113-5: 2011	In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 5: In vitro diagnostic instruments for self-testing
EN ISO 15223-1: 2016	Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements
EN 13532: 2002	General requirements for in vitro diagnostic medical devices for self-testing
EN 13612: 2002	Performance evaluation of in vitro diagnostic medical devices
EN ISO 23640: 2015	In vitro diagnostic medical devices. Evaluation of stability of in vitro diagnostic reagents
EN ISO 14971: 2012	Medical devices - Application of risk management to medical devices
EN ISO 15197: 2015	In vitro diagnostic test systems - Requirements for blood-glucose monitoring systems for self-testing in managing diabetes mellitus
EN ISO 17511: 2003	In vitro diagnostic medical devices - Measurements of quantities in samples of biological origin - Metrological traceability of values assigned to calibrators and control materials
EN 61010-1: 2010 Third edition	Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1 : General requirements
EN 61010-2-101: 2002	Safety requirements for electrical equipment for measurement, control and laboratory use. Particular requirements for in vitro diagnostic (IVD) medical equipment
EN 61326-2-6: 2006	Electrical equipment for measurement, control and laboratory use. EMC requirements. Particular requirements. In vitro diagnostic (IVD) medical equipment
EN 60068-2-64: 2008	Environmental testing. Tests. Test Fh. Vibration, broadband random and guidance
EN 62304: 2006	Medical device software - Software life cycle processes
EN 62366: 2008	Medical devices - Application of usability engineering to medical devices