

EC Declaration of Conformity

We, **Kaz Europe Sàrl**, Place Chauderon 18, 1003 Lausanne, Switzerland declare under our sole responsibility that the product to which this declaration relates is in conformity with the following standard(s) or other normative document(s) and proves the conformity of the designated product with the provisions of the below Directive(s):

- **MDD - Council Directive 93/42/EEC of 14 June 1993 concerning medical devices**
- **RoHS - Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment**

| | |
|---------------------------------------|----------------------|
| Name | Type or model |
| Braun Nasal Aspirator 1 BNA100 series | BNA100EU BNA100KO |

Standards Applied:

| Standard Reference | Edition | Title |
|--------------------|-------------------|---|
| EN 60601-1 | 2006 + A1:2013 | Medical electrical equipment – Part 1: General requirements for basic safety and essential performance. |
| EN 60601-1-11 | 2015 | Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment |
| EN 60601-1-2 | 2015 | Medical Electrical Equipment – Part 1-2: General requirements for Basic Safety and Essential Performance- Collateral Standard: Electromagnetic Compatibility- Requirements and Tests |
| EN 60601-1-6 | 2010 | Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability |
| EN ISO 10993-1 | 2009 + AC:2010 | Biological evaluation of medical devices — Part 1: Evaluation and testing. |
| EN ISO 10993-5 | 2009 | Biological evaluation of medical devices — Part 5: Tests for In Vitro cytotoxicity |
| ISO 10993-10 | 2010 | Biological evaluation of medical devices — Part 10: Tests for Irritation and sensitization |
| EN ISO 14971 | 2019 | Medical devices- Application of risk management to medical devices. |
| EN 15223-1 | 2016 | Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements |
| EN 62366-1 | 2015 | Medical devices — Application of usability engineering to medical devices |
| EN ISO 14155 | 2011 + AC:2011 | Clinical Investigation of medical devices for human subjects - Good clinical practice |
| EN 1041 | 2008 + A1:2013 | Information supplied by the manufacturer with Medical Devices |
| EN 62366-1 | 2015 | Medical devices — Application of usability engineering to medical devices. |
| EN 60601-1-6 | 2010 | Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability |
| EN ISO 14155 | 2011 | Clinical Investigation of medical devices for human subjects - Good clinical practice |

The Technical Documentation is the responsibility of: **Kaz Europe Sàrl**, Place Chauderon 18, CH-1003 Lausanne, Switzerland

EC Declaration of Conformity

Additional Information:

| | |
|--|--|
| For Medical Device Directive 93/42/EC | |
| Regulatory class (MDD, Annex IX): | class IIa (Annex IX rule 11) |
| Conformity assessment procedure: | Annex V |
| GMDN | 56647 |
| UMDNS | 10-216 |
| Notified Body | DQS Medizinprodukte GmbH August Schanz Str. 21 D-60433 Frankfurt, Germany Registration number: 0297 |
| EC Certificate | 381008 MR5 |
| EN ISO 13485 Certificate | 381008 MP2016 |

Authorized Representative in Europe:
Address:

Obelis, S.A.
Bd. Général Wahis, 53
1030 Brussels, Belgium

Authorized Representative in Turkey:
Address:

Sistem Çözüm Ortaklığı Satış Dağıtım Tic. Ltd. Şti.
Ortaklar Cad. Bahçeler Sok.
18 İş Merkezi K:3 D:5 Mecidiyeköy
34394 İstanbul, Turkey
+90 212 216 2950

Tel:

This declaration of conformity is valid until May 26, 2024.

Michael Burke



Lausanne

December 10, 2021

General Manager EMEA

Legally binding signature

Place

Date

Company Stamp:



Kaz Europe Sàrl
Place Chauderon 18
1003 Lausanne
Switzerland
Tel. +41 21 644 01 10
Fax. +41 21 644 01 11