

We, the manufacturer, herewith declare that the products

**Product: Motorola Care 3-1 Non-contact baby thermometer**  
**Model: MBP66NT**

*(including system components and accessories)*  
meet the provisions of Directive 93/42/EEC which apply to them.

The medical device has been assigned to class IIa according to Annex IX Rule 10, of the Directive 93/42/EEC.  
It bears the mark



The product concerned has been designed and manufactured under a quality management system according to Annex III, excluding sec.4, of Directive 93/42/EEC, and the essential requirements of Annex I pertaining to medical devices.

Compliance of the designated product with the Directive 93/42/EEC has been assessed and certified by the Notified Body

**SGS Fimko Ltd**  
**Takamatie 8**  
**FI-00380 Helsinki, Finland**  
**Notified Body 0598**  
**Country: Finland**  
**Certificate No.: FI20/00003**  
**Issue date: 2020-02-23**  
**Expiry date: 2024-03-24**

following the procedure relating to the EC Declaration of Conformity set out in Annex II (excluding sec.4) & Annex VII, of Directive 93/42/EEC.

The above mentioned declaration of conformity is exclusively under the responsibility of

**AVITA Corporation**