

6. Turn off: Press the power button for 3 seconds to turn off the unit.
7. Auto power shut off: The unit will automatically shut off after 14 minutes of any operation for 14 minutes.
8. Change battery: When the symbol "⎓" shown on the LCD, please change the battery.

D. Specification

A. Air velocity				Accuracy
Unit	Range	Resolution	Threshold	
M/s	0~30	0.1	0.1	±5%
Ft/min	0~5860	19	39	
Knots	0~55	0.2	0.1	
Km/hr	0~90	0.3	0.3	
Mph	0~65	0.2	0.2	
B. Temperature				Accuracy
Unit	Range	Resolution		
°C	-10°C~+45°C	0.2		±2°C
°F	14°F~113°F	0.36		±3.6°F
Battery	CR2032 3.0V (Included)			
Thermometer	NTC thermometer			
Operating temperature	-10°C~+45°C (14°F~113°F)			
Operating humidity	Less than 90%RH			
Store temperature	-40°C~+60°C (-40°F~140°F)			
Current consumption	Approx. 3mA			
Weight	52g			
Dimension	40x18x105mm			

Specific Declarations:

Our company shall hold no any responsibility resulting from using output from this product as an direct or indirect evidence. We reserves the right to modify product design and specification without notice.

