

Manufacturer: Beurer GmbH (see address in footer)

SRN: DE-MF-000005422

Product category: Infrared lamp

Product type: IL 60

Intended use: The infrared lamp is intended for irradiating the human body. By irradiating with infrared light, heat is transported to the human body.

Article number: 616.95

The product specified above is in conformity with the following specifications. The products subjected within this declaration of conformity can be identified by the above product type, the CE-mark followed by the four-digit number 0483 and an MD-symbol on the device.

(EU) 2017/745 Medical device regulation (MDR)

Basic-UDI-DI: 4211125IL60PB

Classification/applied rule(s): Class IIa/rule 9

Conformity assessment procedure: Annex IX, Chapter I

The notified body mdc medical device certification GmbH, located at Kriegerstr. 6, 70191 Stuttgart, Germany, identification number 0483, issued in the course of the mentioned conformity assessment procedure the following certificate:

Certificate no. and validity: D1311700055, valid to 2026-04-07

(EU) 2016/425 Personal protective equipment (PPE)

Conformity assessment procedure: Annex V

The notified body ECS European Certification Service GmbH, located at Hüttfeldstraße 50, 73430 Aalen, Germany, identification number 1883, performed the EU type-examination (Module B) and issued the following EU type-examination certificate.

Certificate no. and validity: C3440.1BEU, valid to 2026-12

EN 171:2002
EN 166:2001

2011/65/EU Restrictions of the use of certain hazardous substances in electrical and electronic equipment (RoHS)


EN IEC 63000:2018

This declaration of conformity is issued under the sole responsibility of the manufacturer.

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 2023-09-20

Name, function, signature, stamp: Werner Meternek, Director Quality Management & Regulatory Affairs

ppa.  **Beurer GmbH**
Soeflinger Straße 218 • 89077 Ulm