



### EU DECLARATION OF CONFORMITY

<b>Technical Documentation Name</b>	senofilcon A with Light Filtering Additive contact lenses		
<b>Version Number</b>	2.0		
<b>Product Identification</b>	<b>Trade Name of Device</b> The following product listing includes Diagnostic, Revenue, and Kit Configurations:	<b>Device Name</b>	<b>Basic UDI-DI</b>
	ACUVUE® OASYS MAX 1-Day Contact Lenses	ACUVUE OASYS MAX 1-DAY	0733905a00011BK
	ACUVUE® OASYS MAX 1-Day MULTIFOCAL Contact Lenses		
<b>Legal Manufacturer</b>	Johnson & Johnson Vision Care, Inc. 7500 Centurion Parkway Jacksonville, Florida 32256 United States		
<b>EU Authorised Representative</b>	AMO Ireland Block B, Liffey Valley Office Campus, Quarryvale, Co. Dublin D22 X0Y3, Ireland		
<b>Notified Body</b>	BSI Group The Netherlands B.V. Say Building, John M. Keynesplein 9, 1066 EP Amsterdam Netherlands  Phone : +31 (0)20 346 07 80 Notified Body number : 2797		
<b>Intended Purpose</b>	<p>The ACUVUE® OASYS MAX 1-Day Contact Lenses are intended for Daily Wear for the optical correction of myopia (short-sightedness) and hyperopia (long-sightedness) in persons with healthy eyes that may have 1.00D or less of astigmatism.</p> <p>The ACUVUE® OASYS MAX 1-Day MULTIFOCAL Contact Lenses are intended for Daily Wear for the optical correction of myopia (short-sightedness) and hyperopia (long-sightedness) in presbyopic persons with healthy eyes that may have 0.75D or less of astigmatism.</p> <p>The contact lenses contain a UV blocker to help provide protection against transmission of harmful UV radiation to the cornea and into the eye.</p>		
<b>Classification</b>	IIa		
<b>Product Codes</b>	Universal Product Codes (UPC) for the contact lenses are provided within the ERP (Enterprise Resource Planning) SAP system.		

<b>GMDN Codes</b>	47841, Soft corrective contact lens, daily-disposable
<b>EMDN (CND) Code</b>	Q021004010101, Contact lenses-Hydrogel, Daily Single-Use
<b>Manufacturer's Single Registration Number (SRN)</b>	Not Yet Available
<b>Design, Manufacturing and Distribution Sites</b>	<p>This document is valid for all medical devices described originating from the following sites:</p> <p>Johnson &amp; Johnson Vision Care, Inc. 7500 Centurion Parkway Jacksonville, Florida 32256 United States</p> <p>Johnson &amp; Johnson Vision Care Ireland UC The National Technology Park Limerick V94 N732 Ireland</p>
<b>This Declaration of Conformity is issued under the sole responsibility of the manufacturer.</b>	
We, Johnson & Johnson Vision Care, Inc., hereby declare the above listed medical devices comply with Medical Device Regulation (MDR) 2017/745.	
This declaration is made on the basis of MDR Certificate Number 732087, issued by above stated Notified Body, in accordance with the conformity assessment laid down in Annex IX of MDR 2017/745.	

<b>SIGNATURES</b>		
<b>Place of Issue</b>	<b>Refer to Manufacturer's Address above</b>	
	<b>Date</b>	June 2, 2022
Victoria Brennand Associate Director, Regulatory Affairs Johnson & Johnson Vision Care, Inc. Jacksonville, Florida 32256, USA		
	<b>Date</b>	June 2, 2022
Thomas Wilkinson Director, Quality Systems, Quality Compliance Johnson & Johnson Vision Care, Inc. Jacksonville, Florida 32256, USA		