



Certificate of Compliance

Certificate Number: DL-2019103439C

Applicant: Guangzhou Lesenxinpin Electronic Co., Ltd
3 Floor of Building 3, No.112, longxing west road, Taihe Town, Baiyun District , Guangzhou City, China

Manufacturer: Guangzhou Lesenxinpin Electronic Co., Ltd
3 Floor of Building 3, No.112 , longxing west road, Taihe Town, Baiyun District , Guangzhou City, China

Product: skin care device

Brand Name: lesen

M/N: LS1902
LS1901, LS1903, LS1904, LS1905, LS1906, LS1907, LS1908, LS1909, LS1910

Test Standard: EN 55014-1:2017
EN 61000-3-2:2019, EN 61000-3-3:2013+A1:2019
EN 55014-2:2015
EN 61000-4-2:2009, EN 61000-4-3: 2006+A1:2009+A2:2010
EN 61000-4-4:2012, EN 61000-4-5:2014+A1:2017,
EN 61000-4-6:2014/AC:2015, EN 61000-4-11:2004+A1:2017

The EUT described above has been tested by us with the listed standards and found in compliance with the council EMC directive 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC Directive. It is only valid in connection with the test report number: DL-2019103439E.



This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole product and relevant. Without the written approval, It is not permitted to use the test lab's logo.

Shenzhen DL Testing Technology Co., Ltd.
Part One of 301, A-2 Factory Building, Yalijia Industrial Plant, No. 87, Hengping Road,
Yuanshan Street, Longgang District, Shenzhen, China

Web: www.dl-cert.com E-mail: Service@dl-cert.com Tel: 400-688-3552

