



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND

9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: 407732

Description: Type 2 Sound Meter

Date of Issue: 31-Oct-13

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

EMC Directive: 2004/108/EC

Report Number: WT088000881

Report Date of Issue: 5/6/2008

Standards:

EN 61326-1:2006

EN 61326-2-1:2006

IEC 61326-1:2005

IEC 61326-2-1:2005

The test reports show that the product fulfills the requirement in the EC EMC Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC EMC Directive.



Tony Campagna
Director of Quality Services