



**ISET S.r.l.**

Sede Legale e Uffici  
Via Donatori di sangue, 9 - 46024 Moglia (MN)  
Tel. e fax +39 (0)376 598963  
iset@iset-italia.com [www.iset-italia.eu](http://www.iset-italia.eu)

Cap. soc. i.v. € 10.200,00  
Cod. Fisc. e P.IVA 02 332 750 369  
Reg. Imprese 02 332 750 369  
REA MN 0221098

## CERTIFICATE OF COMPLIANCE

Certificado de Conformidade - Сертификат соответствия - Konformitätserklärung

**1) APPLICANT:**

Ningbo Yinzhou Joan Lab Equipment Co., Ltd.  
No.34 Xiyanshi Road, Yimen Village, Jiangshan Town,  
Yinzhou, Ningbo, Zhejiang, China

**2) CERTIFICATE NO.:**

IT1403YZ10101802

**TEST REPORT(S) NO.:**

TMC180910110-S,  
TMC180910110-E  
TMC Testing Services(Shenzhen) Co., Ltd.

**3) WITH REFERENCE TO EC DIRECTIVE APPLIED:**

Low Voltage Directive 2014/35/EU  
Electromagnetic Compatibility 2014/30/EU

**HARMONIZED STANDARDS APPLIED:**

EN 61010-1: 2010,  
EN 55011:2009+A1:2010,  
EN 61326-1:2013  
EN 61000-3-2:2014,  
EN 61000-3-3:2013

**4) CERTIFICATION ISET MARK:**



**5) PRODUCT CHARACTERISTICS:** Magnetic Stirrer

**MODEL(S):** HS-12,HS-17,HS-19,HSC-19,HSC-19T,OS-40,OS-40C,OS-60C,HM, HMC,HMS,HMSC

**REMARK:** The verification has been carried out on a voluntary basis. We attest that a TCF is in place. The product(s) satisfies the requirements of the Certification Mark of ISET, in reference to the above list standard(s). The above compliance mark can be fixed on the product(s) according to the ISET regulation about its release. This verification doesn't imply assessment of the production and the product(s).



Notice of the CE marking: The label of the CE marking: Not less than 5mm height. Before putting the product(s) into market, CE marking and EC declaration are duties of the manufacturer. The manufacturer is responsible to start the CE marking certification procedure and to perform the activities according to all the relative directive(s).

**6) DATE OF ISSUE:** 10/10/2018

**DATE OF EXPIRE:** 09/10/2021

**CERTIFICATION MANAGER:**

